# **EC CERTIFICATE**

## **Production Quality Assurance**

Certificate No.: 242878-2017-CE-NOR-NA-PS Rev. 1.0

Project No.: PRJC-102229-2008-MSL-NOR

Valid Until: 27 May 2024

This is to certify that the quality system of:

## **Laerdal Medical AS**

P.O. Box 377, Tanke Svilandsgate 30, 4002 Stavanger, Norway

For production and final product inspection/testing of:

## **Disposable Resuscitators**

Has been assessed with respect to:

THE CONFORMITY ASSESSMENT PROCEDURE DESCRIBED IN ANNEX V OF COUNCIL DIRECTIVE 93/42/EEC ON MEDICAL DEVICES, AS AMENDED

and found to comply.

Further details of the product(s) and conditions for certification are given overleaf.

Place and date: Høvik, 10 February 2020



PROD 021

For:
DNV GL PRESAFE AS
Notified Body No.: 2460

Cathrine Wisbech

The certificate is digitally verified by blockchain technology. For more info, see www.dnvgl.com/assurance/certificates-in-the-blockchain.html



Notice: The Certificate is subject to terms and conditions as set out in the Certification Agreement. Failure to comply may render this Certificate invalid.

MSD-CO-078-B Rev 0.0 Page 1 of 3



Certificate No.: 242878-2017-CE-NOR-NA-PS Rev. 1.0

Project No.: PRJC-102229-2008-MSL-NOR Valid Until: 27 May 2024

#### Jurisdiction

Application of Council Directive 93/42/EEC of 14 June 1993, adopted as "Forskrift om Medisinsk Utstyr" by the Norwegian Ministry of Health and Care Services.

#### Certificate history:

Revision	Description	Issue Date
0.0	Replaces DNV GL (NB 0434) certificate no. 158818-2014- CE-NOR-NA ver 1.0 following transfer of Notified Body functions to DNV GL Nemko Presafe AS (NB 2460).	16 November 2017
1.0	Recertification	10 February 2020

#### Products covered by this Certificate:

<b>Product Description</b>	Product Name	Class
Disposable Resuscitators	The Bag II Disposable Resuscitator – Adult, Child and Infant Configurations	IIa
Accessory	Laerdal Disposable Peep Valve Assembly (5-20 cm H2O) Laerdal Disposable Masks with inflation port	IIa

The complete list of devices is filed with the Notified Body

#### Sites covered by this certificate

Site Name	Address
Postal Address: Laerdal Medical AS	P.O. Box 377, Tanke Svilandsgate 30, 4002 Stavanger, Norway
Physical Address: Laerdal Medical AS	Tanke Svilandsgate 30, 4007 Stavanger, Norway



Certificate No.: 242878-2017-CE-NOR-NA-PS Rev. 1.0

Project No.: PRJC-102229-2008-MSL-NOR

Valid Until: 27 May 2024

#### **Terms and conditions**

The certificate is subject to the following terms and conditions:

- Any producer (see 2001/95/EC for a precise definition) is liable for damage caused by a defect in his product(s), in accordance with directive 85/374/EEC, as amended, concerning liability of defective products.
- The certificate is only valid for the products and/or manufacturing premises listed above.
- The Manufacturer shall fulfil the obligations arising out of the quality system as approved and uphold it so that it remains adequate and efficient.
- The Manufacturer shall inform Presafe of any intended updating of the quality system and Presafe will assess the changes and decide if the certificate remains valid.
- Periodical audits will be held, in order to verify that the Manufacturer maintains and applies the quality system. Presafe reserves the right, on a spot basis or based on suspicion, to pay unannounced visits.

The following may render this Certificate invalid:

- Changes in the quality system affecting production.
- Periodical audits not held within the allowed time window.

## Conformity declaration and marking of product

When meeting with the terms and conditions above, the producer may draw up an EC declaration of conformity and legally affix the CE mark followed by the Notified Body identification number of Presafe.

End of Certificate

NOTIFIED BODY: DNV GL PRESAFE AS, Veritasveien 3, N-1363 Høvik, Norway - Registered Enterprise No: NO 997 067 401 MVA .

MSD-CO-078-B Rev 0.0 Page 3 of 3